

Dear Competent Authority,

On behalf of Bohus BioTech I need to inform you the following:

URGENT FIELD SAFETY NOTICE

Date: May 17, 2022

Attention: Competent Authority

We have identified deviations in our sterilisation process and therefore we can not guarantee the sterility of the following products.

1. Microvisc 1%
2. Microvisc Plus 1,4%
3. Microvisc Phaco 2,5%

Above include all syringe sizes, 0,55 ml, 0,85 ml and 1,0 ml. This also include all products being under own label/private label.

Therefor we have kindly asked our distributors to immediately take the following actions:

- Recall all products within expiry date.
- This notice needs to be passed on to all those who need to be aware.
- Confirmation of product recall to Bohus BioTech.
- Above needs to be executed as soon as possible.

Magnus Nylén, CEO

Bohus BioTech



Michelle Johansson, Quality Director

Bohus BioTech

